

Test Report.

Report NO. : National textile committee character NO YJ202003582

Name Of Sample: Daily protective mask

Customer: Yushan Kuster Medical Technology Co., Ltd

Test Category: Entrusted inspection

Zhejiang Light Industrial Products Inspection and Research Institute

National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

Zhejiang Light Industrial Products Inspection and Research Institute
National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)

Inspection Report

National textile committee character NO. YJ202003582

Name of Customer	Yushan Kuster Medical Technology Co., Ltd.	Address	---
Manufacturer	---	Address	---
Sample information	Name of sample: Daily protective mask Characteristics: Blue Trademark: --- Specification/model: --- Level: --- Category of safety specification: --- Art. No.: --- -----		
(Above-mentioned information by Customer-supplied)			
The sent way of sample	Express	Sample quantity	2 packs
Receiving Date of Sample	2020-03-17	Test Category	Entrusted inspection
Rating Requirements	GB/T 32610-2016		
Test Summary: <div style="text-align: right;"> Test Seal Date Of Approval: 2020.3.17 </div>			
Remarks	1. The sample identification is not marked with the level of protection effect, which is determined according to the minimum standard D level. 2. The sample is not pretreated		

Approved by

Inspection Report

National textile committee character NO. YJ202003582

No.	Test item		Detection method	Unit	Standard requirements (D)	Measured value	Individual evaluation	Results remarks
1	Respiratory resistance	Inspiratory resistance	GB/T 32610-2016	Pa	≤ 175	52	Consistent	---
		Expiratory resistance		Pa	≤ 145	52		
2	Filter property (Saline medium)		GB/T 32610-2016 Appendix A	%	≥ 90	91.8	Consistent	---
3	Protective effect		GB/T 32610-2016 Appendix B	%	≥ 65	65.0	Consistent	---

Inspection Report

National textile committee character NO. YJ202003582

Sample Photo



DECLARATION

1. Our organization guarantees impartiality, independence and honesty of inspection, and is responsible for the results of inspection, keeping the samples supplied by the entrusting party confidential and at the same time protecting the ownership of the samples supplied.
2. The test report will be deemed invalid without signatures of the inspector/reviewer and authorized personnel, and the red special inspection stamp of our organization.
3. The test report will be invalid if it is altered. Copies of the report are invalid without the red special inspection stamp of our organization.
4. The test results shown in this report is only valid for the tested sample.
5. All the pages of the report are integral parts of the report. Our organization will not be responsible for any misunderstanding or other results caused by using separate page(s) of the report.
6. If there is any dissent of the report, the entrusting party shall notify our organization timely. For the mandatory inspection given by governmental administration departments, and dissent about the sample being tested or test results on the report should be dealt with in accordance with national regulations.

Zhejiang Light Industrial Products Inspection and Research Institute The Affiliated Inspection Centers

Zhejiang Light Industrial Products Inspection and Research Institute
National Textiles and Garment Quality Supervision Inspection Center(Zhejiang)
Address: Building No.6, 300 XiaSha Road, Hangzhou, Zhejiang
Telephone: 0571-85122669
E-Mail: 1137907994@qq.com

National Center for Quality Supervision Inspection of Furniture(Zhejiang)
Zhejiang Center of Quality Test for Indoor Safety and Furniture Products
Address: Building 4 LiangZhu University Science and Technology Park, No.1, JingYi Rd, Yuhang District,
Hangzhou, Zhejiang
Telephone: 0571-89009556
E-Mail: 2047699564@qq.com

Zhejiang Center of Quality Test for Light Industry and Hardware Products
Zhejiang Center of Quality Test for Sports Products
Address: No. 222 Tianmushan Rd., Hangzhou, Zhejiang Building 3 LiangZhu University Science and Technology
Park, No.1, JingYi Rd., Hangzhou, Zhejiang
Telephone: 0571-89001107
E-Mail: wujtest@126.com

National Center for Quality Supervision Inspection of Lock(Zhejiang)
Zhejiang Center of Quality Test for Lock Products
Address: No. 24 Tangmiao Rd., Hangzhou, Zhejiang
Telephone: 0571-85027738
E-Mail: Locktest@126.com