

Test Confirmation

BINDER
Best conditions for your success

Model: <u>115</u>	ED ☒ , FD ☐ , FED ☐ , FP ☐ , M ☐ , MDL ☐
Series number:	<u>14-15003</u>

We herewith confirm that the above mentioned apparatus has been calibrated in accordance with the certified system for quality assurance DIN EN ISO 9001 (TÜV Reg. certificate no. 12 100 27534 TMS)

Conditions of testing:

The tests have been carried out in accordance with the BINDER instructions for testing. During the tests the apparatus was empty except from two shelves.

Temperature adjustment:


Measuring point: center of the inner chamber with a probe PT 100 / humidity probe
Test conditions: Test temperature: 150 °C / 302 °F

Environmental conditions

Temperature 25 °C / 77 °F ± 5 °C / 44.6 °F
Humidity 50 % RH ± 20 % RH

Type plate checked

Accessories checked see attached packing list



Manager Final Inspection / Date

BINDER

Packing list for ED, FD, FED, FP, M, MDL

name of component / Bauteil	Quantity	Checked
Operating manual (set) / Betriebsanleitung (Set)	1	☒
Rack chrome plated / Einschubgitter verchromt	2	☒
Safety cover Class 2 / MDL: Schutzabdeckung Klasse 2	1	☐

BINDER GmbH | Im Mittleren Ösch 5 | D-78532 Tuttlingen

Tel. + 49 (0) 7462 / 2005 – 0 | Fax: + 49 (0) 7462 / 2005 – 100 | www.binder-world.com

500.05.001

Autor: JLE

Geprüft: JBR

Freigegeben: JBR

Datum: 31.01.12

Index: 00